## Intoxilyzer® 8000



All new infrared spectrometry mobile, evidential breath alcohol instrument

- Uses both 3 and 9 micron technology
- Fully mobile or stationary use
- Easy single button operation
- Rugged ABS case



## A NEW INSTRUMENT FROM THE LEADER IN EVIDENTIAL DUI ENFORCEMENT

Intoxilyzer.

THE ALL-NEW INTOXILYZER® 8000 COMES FROM CMI, INC., THE MANUFACTURER OF THE LEGENDARY INTOXILYZER® 5000 SERIES AND ITS TRADITION OF CASE LAW-APPROVED INFRARED SPECTROMETRY ALCOHOLANALYSIS. THE 8000 TAKES ACCURATE AND RELIABLE MEASUREMENT TO A NEW LEVEL BY USING BOTH 3 AND 9 MICRON TECHNOLOGY...TOGETHER!



The integral keyboard detaches for easy use and test subject data entry.



The rear panel includes three (two visible) RS-232 ports, printer port and an internal modem.

## SPECIFICATIONS

- ANALYTICAL PRINCIPLE: infrared absorption, at both 3 and 9 microns on every sample
- RANGE: 0.000 to 0.600 %BAC [g/210 Liters]; 0.000 to 3.000 mg/liter.
- DIMENSIONS: 14.5" long x 10" wide x 8.4" deep.
- WEIGHT: 17 lbs [instrument only].
- CALIBRATION CHECKS: wet bath simulator or dry gas.
- BAROMETRIC COMPENSATION: automatic compensation during dry-gas calibration checks.
- **DISPLAY:** 2 line by 20 character, 8mm character height vacuum fluorescent display.
- PRINTER OPTIONS: thermal, impact, external
- AUDIBLE BEEPER: gives auxiliary messages: tones are software-variable in frequency and duration.
- POWER RATINGS: AC: 90 – 264 VAC, 47 – 63Hz, 1amp max. Vehicle DC: 10 – 16 VDC, 6 amps max.
- MICROPROCESSOR: Intel compatible x86 based processor @ 30 MHz with a separate DSP processor for signal conditioning.

- MEMORY: 1 Mbyte for data storage [expandable to 4 Mbyte]; 1 Mbyte FLASH/SRAM for code/fast access. Remote program updates.
- REAL-TIME CLOCK: accurate to ±10 minutes per year.
- WARM-UP TIME: less than 20 minutes at 20°C
- **TEST TIME:** less than 1 minute, excluding data entry.
- DISPOSABLE MOUTHPIECES: spit-trap individually wrapped
- COMMUNICATIONS: Three RS-232 serial ports, one external printer port, internal modem
- CE CERTIFICATION: complies with IEC 1010-1 (1990, 92, 95)
- OIML: complies with OIML R 126 [1998 (E)]
- RFI CERTIFICATION: to BS EN 50081/2-1 97 & EC 95/54/ EEC and FCC Part 15 Class B.
- VIBRATION AND SHOCK: to IEC 68-1/2.
- CASE CONSTRUCTION: injection molded ABS.
- SOFTWARE CONFIGURATION: various software settings may be configured by the instrument *Manager* through the keyboard [via a security password] / through an external PC using COBRA.

